

# Haemoglobin (Ready-To-Use) Drabkin's Solution

REF:611 001 (1 x 250 ml) 100 test  
REF:611 002 (2 x 250 ml) 200 test  
REF:611 003 (2 x 500 ml) 400 test

## Intended Use

Spectrum Diagnostics haemoglobin reagent is intended for the in-vitro quantitative, diagnostic determination of haemoglobin in human blood.

## Background

Haemoglobin (Hb) is the red pigmented protein located in the erythrocytes and consists of four subunits. Its main function is the transport of oxygen and carbon dioxide in blood. In normal humans, at least 96 % of the haemoglobin is HbA. HbA2 is usually about 2.5 – 3 % of total haemoglobin. Fetal hemoglobin (HbF) predominates during fetal life and diminishes rapidly during the first year of postnatal life. In normal adults less than 1 % is HbF. Blood haemoglobin concentration may be diminished as a consequence of hemorrhage or hemolysis or as a result of impaired blood formation in the bone marrow.

## Method

colorimetric method using Drabkin's solution.

## Assay Principle

Haemoglobin is oxidized by potassium ferricyanide which is converted into stable cyanomethaemoglobin by potassium cyanide. The absorbance of the cyanomethaemoglobin is monitored at 540 nm.

## Reagents

### Reagent

Potassium ferricyanide	0.62 mmol/l
Potassium phosphate	1.04 mmol/l
Potassium cyanide	1.54 mmol/l
Surfactant	< 0.1 %

**Harmful (Xn): R20/21/22:** Harmful by inhalation, in contact with skin and if swallowed. S7: Keep container tightly closed. S28.1: After contact with skin, wash immediately with plenty of water.

**S45:** In case of accident or if you feel unwell, seek medical advice immediately. The amount of cyanide present in one bottle of reagent is appreciably less than the minimum lethal dose for an adult. However, hydrogen cyanide is liberated by acidification. Never allow reagent to come in contact with acid.

For further information, refer to the Haemoglobin reagent material safety data sheet.

## Precautions and Warnings

Do not ingest or inhale. In case of contact with eyes or skin; rinse immediately with plenty of soap and water. In case of severe injuries; seek medical advice immediately.

## Reagent Preparation Storage and Stability

Reagent is supplied ready to Use.

Reagent is stable until expiration date stated on label when stored at 15 - 25 °C. Once opened, the opened vial is stable for 6 months at the specified temperature.

## SYMBOLS IN PRODUCT LABELLING

	Authorised Representative		Use by/Expiration Date
	For in-vitro diagnostic use		CAUTION. Consult instructions for use
	Batch Code/Lot number		Manufactured by
	Catalogue Number		(Xn) - Harmful
	Consult instructions for use		Temperature Limitation

## Specimen Collection and Preservation

Anticoagulated venous or capillary blood. Blood may be anticoagulated with EDTA, or fluoride. Blood can be taken directly from a finger or heel puncture without use of anticoagulant.

**Stability :** 7 days at 2 – 8 °C  
4 days at 20 - 25 °C

## System Parameters

Wavelength	540 nm (Hg 546 nm)
Optical path	1 cm
Assay type	End-point
Direction	Increase
Sample : Reagent Ratio	1 : 250
e.g. : Reagent volume	2.5 ml
Sample volume	10 µl
Temperature	20 – 25 °C
Incubation time	5 minutes
Zero adjustment	Against reagent blank

## Procedure

### Pipette into test tubes

Reagent	2.5 ml
Blood sample	10 µl

Mix well and rinse the blood pipette several times with the reagents, and incubate for 5 minutes at 20-25 °C. Measure absorbance of specimen (Aspecimen) against reagent blank.

## Calculation

Haemoglobin concentration (g/dL) = Aspecimen x 36.77

Haemoglobin concentration (mmol/L) = Aspecimen x 22.83

## Quality Control

Normal & abnormal commercial control blood of known concentrations should be analyzed with each run.

## Performance Characteristics

### Precision

Within run (Repeatability)

	Level 1	Level 2
n	20	20
Mean (g/dL)	10	14
CV%	2.3	1.3

Run to run (Reproducibility)

	Level 1	Level 2
n	20	20
Mean (g/dL)	11.1	14.1
CV%	2.9	2.1

## Methods Comparison

A comparison between Spectrum Diagnostics Haemoglobin reagent and a commercial reagent of the same methodology was performed on 40 human blood samples. A correlation (r) of 0.983 was obtained.

## Expected values

1- 6 days	15.2 – 23.5 g/dL	( 9.4 – 14.6 mmol/L)
14 – 50 days	10.3 – 16.6 g/dL	( 6.4 – 10.3 mmol/L)
2 - 10 months	10.0 – 12.9 g/dL	( 6.1 – 8.0 mmol/L)
1 – 15 years	11.0 – 14.3 g/dL	( 6.8 – 8.8 mmol/L)
Adults Women	12.0 – 16.0 g/dL	(7.5 – 9.9 mmol/L)
Men	14.0 – 18.0 g/dL	(8.7 – 11.2 mmol/L)

Spectrum Diagnostics does not interpret the results of a clinical laboratory procedure; interpretation of the results is considered the responsibility of qualified medical personnel. All indications of clinical significance are supported by literature references.

## Waste Disposal

This product is made to be used in professional laboratories. Please consult local regulations for a correct waste disposal.

**S56:** dispose of this material and its container at hazardous or special waste collection point.

**S57:** use appropriate container to avoid environmental contamination.

**S61:** avoid release in environment. refer to special instructions/safety data sheets.

## References

1. International committee for standardization in haematology. Brit. J. Haemat., 1967:13 (Suppl.) 71.
2. Van Kampen, E. J. and Zijlstra, W.G., Clin. Chem. Acta., 1961:6:538 – 544.
3. Tietz NW, Ed. Clinical guide to laboratory tests. 2ND ED. Philadelphia: WB Saunders; 1990:566.

## ORDERING INFORMATION

CATALOG NO.	QUANTITY
611 001	1 x 250 ml
611 002	2 x 250 ml
611 003	2 x 500 ml



### Egyptian Company for Biotechnology (S.A.E)

Obour city industrial area. block 20008 piece 19 A. Cairo. Egypt.

Tel: +202 4489 2248 - Fax: +202 4489 2247

www.spectrum-diagnostics.com

E-mail: info@spectrum-diagnostics.com



**MDSS GmbH**  
Schiffgraben 41  
30175 Hannover, Germany



IFUFHR08

Rev.(3), 19/11/2019