

Lactate dehydrogenase (LDH)-Liquizyme (4+1) E.C.1.1.1.27.

REF: 283 001 (4 x 20 ml) 80 test
 REF: 283 002 (10 x 10 ml) 100 test
 REF: 283 003 (5 x 20 ml) 100 test
 REF: 283 004 (9 x 20 ml) 180 test
 REF: 283 005 (4 x 60 ml) 240 test

Intended Use

Spectrum Diagnostics liquizyme LDH reagent is intended for the in-vitro quantitative, diagnostic determination of LDH in human serum on both automated and manual systems.

Background

The lactate dehydrogenase (LDH) enzyme is widely distributed in heart, liver, muscle, and kidney. LDH catalyzes the conversion of lactate to pyruvate. The enzyme is a tetrameric protein and gives rise to five isoenzymes. Heart, kidney, brain and erythrocytes have the highest proportion of LD-1 and LD-2. Liver and skeletal muscle have highest percentage of LD-5. LDH is significantly increased during myocardial infarction. A maximum value is reached 48 hours after the onset of manifestation and persists up to 10 days. Elevated serum levels of LDH have also been observed in patients with megaloblastic anemia, disseminated carcinoma, leukemia, and trauma. Mild increases in LDH activity has been reported in cases of haemolytic anemia, muscular dystrophy, pulmonary infarction, hepatitis, nephrotic syndrome, and cirrhosis.

Method

Kinetic UV method.

Assay Principle

LDH catalyzes the reaction between pyruvate and NADH to produce NAD⁺ and L-Lactate:



The initial rate of the NADH oxidation is directly proportional to the catalytic LDH activity. It is determined by measuring the decrease in absorbance at 340 nm.

Reagents

Reagent 1 (R1 Buffer)

Phosphate buffer (pH 7.5) 50 mmol/L
 Pyruvate 3.0 mmol/L
 Sodium Azide 8.0 mmol/L

Reagent 2 (R2 Coenzyme)

NADH > 0.18 mmol/L
 Sodium azide 8.0 mmol/L

For further information, refer to the Lactate dehydrogenase reagent material safety data sheet.

Precautions and Warnings

Do not ingest or inhale. In case of contact with eyes or skin; rinse immediately with plenty of soap and water. In case of severe injuries; seek medical advice immediately.

Both reagents (R1) and (R2) contain sodium azide which may react with copper or lead plumbing.

Reagent Preparation

REF:283 001: add 4 ml from R2 to one bottle of R1; mix gently.
 REF:283 002: add 2 ml from R2 to one bottle of R1; mix gently.
 REF:283 003: add 4 ml from R2 to one bottle of R1; mix gently.
 REF:283 004: add 4 ml from R2 to one bottle of R1; mix gently.
 REF:283 005: add one bottle of R2 to one bottle of R1; mix gently.

Or prepare the working solution according to the number of tests required by mixing 4 volumes of reagent 1 (R1) and 1 volume of reagent 2 (R2), e.g. 400 µl R1 + 100 µl R2.

SYMBOLS IN PRODUCT LABELLING

	Authorised Representative		Temperature Limitation
	For in-vitro diagnostic use		Use by/Expiration Date
	Batch Code/Lot number		CAUTION. Consult instructions for use
	Catalogue Number		Manufactured by
	Consult instructions for use		

Reagent Storage and Stability

All reagents are stable until expiration date stated on label when stored refrigerated at 2 - 8 °C.
 Working solution is stable for 3 weeks at 2 - 8 °C or 2 days at 15 -25 °C.

Deterioration

Do not use liquizyme LDH reagent if it is turbid or if the absorbance of the working reagent is less than 1.0 at 340 nm.
 Failure to recover control values within the assigned range may be an indication of reagent deterioration.

Specimen Collection and Preservation

Use nonhaemolyzed serum. Heparin is the only acceptable anticoagulant. Sodium citrate and EDTA have an inhibitor effect and must not be used. The biological half-life of LDH in serum is 10 - 54 hours.

Stability: 6 weeks at 4 - 8°C ; 4 days at 20 - 25°C
 Freezing of the samples is not recommended.

System Parameters

Wavelength	340 nm (334 - 365 nm)
Optical path	1 cm
Assay type	Kinetic
Direction	decrease
Sample : Reagent Ratio	1 : 50
e.g.: Reagent volume	1 ml
Sample volume	20 µl
Temperature	37 °C
Equilibration time	30 seconds.
Read time	1 to 3 minutes
Zero adjustment	Against air
Reagent Blank Limits	Low 1.00 AU High 2.5 AU
Sensitivity	10 U/L
Linearity	1200 U/L

Procedure

Pipette into cuvette (37 °C):	Macro	Semi-Micro
Working solution	1 ml	500 µl
Specimen	20 µl	10 µl

Mix, read initial absorbance after 30 seconds. and start timer simultaneously. Read again after 1, 2 and 3 minutes. Determine the mean absorbance change per minute (ΔA/min).

Calculation

To calculate the LDH activity use the following formula
 U/L = 8095 x ΔA 340 nm/min.
 U/L = 15000 x ΔA 365 nm/min.

Quality Control

Normal & abnormal control serum of known concentrations should be analyzed with each run.

Performance Characteristics

Precision

Within run (Repeatability)

	Level 1	Level 2
n	20	20
Mean (U/L)	433	923
SD	6.8	6.64
CV%	1.57	0.71

Run to run (Reproducibility)

	Level 1	Level 2
n	20	20
Mean (U/L)	439	935
SD	7.1	6.71
CV%	1.62	0.79

Methods Comparison

A comparison between Spectrum Diagnostics LDH reagent and a commercial reagent of the same methodology was performed on 20 human sera. A correlation of 0.967 was obtained.

Sensitivity

When run as recommended, the minimum detection limit of this assay is 10 U/L.

Linearity

The reaction is linear up to LDH concentration of 1200 U/L; specimens showing higher concentration should be diluted 1+5 with physiological saline and repeat the assay (resultx6).

Interfering substances

Serum, plasma

Haemolysis

Erythrocyte contamination elevates results significantly since LDH activities in erythrocytes are 150 times higher than those in normal sera.

Icterus

No significant interference.

Lipemia

Lipemic specimens may cause high absorbance flagging. Diluted sample may be recommended.

Anticoagulants

EDTA and citrate may inhibit the reaction.

Expected values (at 37 °C)

Adults at 240-480 U/L (4.0- 8.0 µkat/L)

Children (7-12 Years)

Female : < 580 U/L (< 9.65 µkat/L)

Male : < 764 U/L (< 12.7 µkat/L)

Premature : < 1103 U/L (< 18.4 µkat/L)

Calculate for temperature conversion factor of 0.5 (37 → 25°C) and 0.67 (37 → 30°C).

Spectrum Diagnostics does not interpret the results of a clinical laboratory procedure ; interpretation of the results is considered the responsibility of qualified medical personnel. All indications of clinical significance are supported by literature references.

Analytical Range:

10 - 1200 U/L.

Waste Disposal

This product is made to be used in professional laboratories. Please consult local regulations for a correct waste disposal.

S56: dispose of this material and its container at hazardous or special waste collection point.

S57: use appropriate container to avoid environmental contamination.

S61: avoid release in environment. refer to special instructions/safety data sheets.

References

- 1.Dito WR. Lactate dehydrogenase: A brief review. In: Griffiths JC, ed. Clinical Enzymology. New York :masson publishing USA; 1979:18
- 2.Kachmar JF, Moss DW: Enzymes. In Fundamentals of clinical chemistry. NW Tietz, editor, saunders, philadelphia, 1976 pp 652-6603.
- 3.Van der heiden C, B Ais, Gerh Ardt W,Rosallsis. Approved recommendation on IFCC methods for the measurement of catalytic concentration of enzymes. Part 8. IFCC method for LDH.Eur J Clinical Chem Clin Biochem. 1994;32:639-655
4. young DS.Effects of drugs on clinical laboratory tests. AACC press, Washington D.C., 1990
5. Zimmerman HJ, henery JB: Clinical enzymology. In: Clinical diagnosis and management by laboratory methods, 16 th., JB Henery, editor, saunders, philadelphia,1979, pp 365-368.

ORDERING INFORMATION

CATALOG NO.	QUANTITY
283 001	4 x 20 ml
283 002	10 x 10 ml
283 003	5 x 20 ml
283 004	9 x 20 ml
283 005	4 x 60 ml



Egyptian Company for Biotechnology (S.A.E)

Obour city industrial area. block 20008 piece 19 A. Cairo. Egypt.

Tel: +202 4489 2248 - Fax: +202 4489 2247

www.spectrum-diagnostics.com

E-mail:info@spectrum-diagnostics.com



MDSS GmbH
Schiffgraben 41
30175 Hannover, Germany



IFUFCC76

Rev.(4), 1/2/2019