SPECTRUM

The Creative Approach to Bioscience

Potassium
Tetraphenylborate Method
Without Deproteinization

REF: 298 001 (2 x 25ml) 50 test
REF: 298 002 (4 x 25ml) 100 test
REF: 298 003 (2 x 100ml) 200 test
REF: 298 004 (4 x 100ml) 400 test

SYMBOLS IN PRODUCT LABELLING

EC REP	Authorised Representative		Use by/Expiration Date
IVD	For in-vitro diagnostic use		CAUTION. Consult instructions for use
LOT	Batch Code/Lot number		
REF	Catalogue Number		Manufactured by
	Consult instructions for use		(Xi) - Irritant
	Temperature Limitation		

Intended Use

Spectrum-Diagnostics Potassium reagent is intended for the in-vitro quantitative diagnostic estimation of potassium in human serum or Plasma on manual systems.

Background

Sodium and Potassium are the major cations of extracellular and intracellular fluids respectively. Sodium maintains the normal distribution of water and the osmotic pressure in the various fluid compartments. Potassium influences the acid base balance and osmotic pressure including water retention. Increased sodium levels are found in severe dehydration and excessive treatment with sodium salts. Decreased levels are found in severe polyurea, metabolic acidosis, diarrhea and renal insufficiency. Increased potassium levels are found in renal failure, dehydration, shock and adrenal insufficiency. Decreased levels are found in malnutrition, gastrointestinal fluid loss, and hyperactivity of the adrenal cortex.

Method

Turbidimetric Tetraphenylborate (TPB)

Assay Principle

At an alkaline pH Potassium ions and TPB form a turbid emulsion, the increase of which can be measured quantitatively in a photometer at 578 nm. The increase of the absorbance (A) is directly proportional to the concentration of Potassium in the sample.

Reagent

Reagent R	NaOH	0.50 mol/L
	TPB-Na	240 mmol/L

Irritant (Xi): R36/38: Irritating to eyes and skin. S26: In case of contact with eyes, rinse immediately with plenty of water and seek medical advice. S37/39: Wear suitable gloves and eye/face protection.

Standard	Potassium	5.00 mmol/L

Precautions and Warnings

Do not ingest or inhalate. In case of contact with eyes or skin; rinse immediately with plenty of soap and water. In case of severe injuries; seek medical advice immediately.

Reagent Storage and Stability

Reagent and standard are ready-to-use. When stored at 2 – 8 ^{O}C; they are stable up to the expiry date printed on the label. The remaining stability after opening the bottles is 1 month at 2 – 8 ^{O}C

Samples

Human Serum .

System Parameters

Wavelength	578 nm
Optical path	1 cm
Assay type	Colorimetric end-point
Direction	Increase
Sample: Reagent Ratio	1:50
e.g.: Reagent volume	1 ml
Sample volume	20 µl
Temperature	25-37 ^{O}C
Equilibration Time	30 seconds
Zero adjustment	Against reagent blank
Linearity	10 mmol/L

Procedure

	Reagent Blank	Standard	Sample
Reagent R	1mL	1 mL	1 mL
Standard	.....	20 µL	.....
Sample	.....	.....	20 µL

Mix, incubate for 3 minutes at 37 ^{O}C or 5 minutes at 25 ^{O}C, Mix again thoroughly and read absorbance of sample (A_{sample}) and standard ($A_{standard}$) against blank.

Calculation

$$\text{Serum Potassium Conc.(mmol/L)} = \frac{A_{Sample}}{A_{Standard}} \times 5$$

Expected Values

Serum	3.6 - 5.5 mmol/L
Plasma	4.0 – 4.8 mmol/L

Note:
It is recommended for each laboratory to establish and maintain its own reference values. The given data are only an indication.

Spectrum Diagnostics does not interpret the results of a clinical laboratory procedure; interpretation of the results is considered the responsibility of qualified medical personnel. All indications of clinical significance are supported by literature references.

Sensitivity

When run as recommended, the minimum detection limit of the assay is 1.5 mmol/L.

Linearity

The assay is linear up to 10 mmol/L

Interfering substances

Interferences are found according to the literature.

Waste Disposal

This product is made to be used in professional laboratories.
Please consult local regulations for a correct waste disposal.
S56: dispose of this material and its container at hazardous or special waste collection point.
S57: use appropriate container to avoid environmental contamination.
S61: avoid release in environment. refer to special instructions/safety data sheets.

References

1. Hillman, G.; Beyer, G.: Z. Klin. Biochem. 5 (1967), 93
2. Hoeflmayr, J.: Praxis und Helferin 8 (1979)
3. Tietz, N.W.: Fundamentals of Clin. Chem. (1976), 876

ORDERING INFORMATION	
CATALOG NO.	**QUANTITY**
298 001	2 x 25 ml 50 Test
298 002	4 x 25 ml 100 Test
298 003	2 x 100 ml 200 Test
298 004	4 x 100 ml 400 Test

Egyptian Company for Biotechnology (S.A.E)
Obour city industrial area. block 20008 piece 19 A. Cairo. Egypt.
Tel: +202 4489 2248 - Fax: +202 4489 2247
www.spectrum-diagnostics.com
E-mail:info@spectrum-diagnostics.com

EC REP **MDSS GmbH**
Schiffgraben 41
30175 Hannover, Germany

ISO 13485:2016