

Hemoglobin A1c (HbA1c) Control Set

REF: 358 001 (2 Levels)

Low Control 1 x 0.5 ml
High Control 1 x 0.5 ml

Intended Use

Accuracy control for quantitative determination of HbA1c in human blood by turbidimetric immunoassay.

composition

The HbA1c controls (Low & High) are hemolysates prepared from packaged human erythrocytes, lyophilized and stabilized.

Preparation

Accurately add 0.5 ml of distilled water to dissolve the contents, Gently mix for 10 minutes, or until all material has dissolved.

Storage and Stability

The expiry date of the product at +2 to +8 °C is listed on the label.
The reconstituted control set is stable for 15 days if stored tightly closed at +2 to +8 °C

Procedure

The controls should be treated in the same manner as patient samples regarding the hemolysate procedure. Follow the directions that accompany the instruments and reagent kit used in the assay.










Precautions and Warnings

1. For "In Vitro Diagnostic Use"
2. All blood donor units comprising the source material used in the manufacture of this product have been tested by approved FDA methods and found to be non-reactive for HBsAg, HCV and anti-HIV 1 & 2.
Only nevertheless every product obtained from human body fluids should be handled with appropriate care in accordance with recommended procedures for biohazardous materials since absence of infectious agents can never be proven.

Assigned Values

The values are lot dependent. Actual values are printed on the labels.

SYMBOLS IN PRODUCT LABELLING

	Authorised Representative		Temperature Limitation
	For in-vitro diagnostic use		Use by/Expiration Date
	Batch Code/Lot number		CAUTION. Consult instructions for use
	Catalogue Number		Manufactured by
	Consult instructions for use		



Egyptian Company for Biotechnology (S.A.E)

Obour city industrial area, block 20008 piece 19 A, Cairo, Egypt

Tel: +202 4665 1848 - Fax: +202 4665 1847

www.spectrum-diagnostics.com

E-mail: info@spectrum-diagnostics.com



MDSS GmbH
Schiffgraben 41
30175 Hannover, Germany



IFUFTI60

Rev.(1), 1/11/2014

