

## Protein standard High

REF:604 001 1x1 ml  
REF:604 002 2x1 ml

### Intended Use

Accuracy standard for the determination of proteins in serum by turbidimetry and nephelometry.

### composition

A dilution of human plasma with phosphate buffered saline.  
Contains stabilizers and 0.095 g% sodium azide .

### Storage and Stability

The expiry date of the product at +2 to +8 °C is listed on the label.  
After first opening the standard can be used for 6 weeks if stored tightly closed at +2 to +8 °C after use.  
Do not freeze.










### Precautions and Warnings

1. For "In Vitro Diagnostic Use"
2. Each individual donation intended for use in manufacturing of protein control serum was tested for hepatitis B surface antigen (HBsAg), anti-hepatitis C virus (anti-HCV) and anti-HIV1 and HIV2 by FDA required test. Only donations with negative findings were used for its manufacture. Nevertheless every product obtained from human body fluids should be handled with appropriate care in accordance with recommended procedures for biohazardous materials since absence of infectious agents can never be proven .
3. Reagents containing sodium azide must be handled with due caution: Do not ingest or allow to contact skin or mucous membranes !

### Assigned Values

The values are lot dependent. Actual values are printed on the Assay Value Sheet .

#### SYMBOLS IN PRODUCT LABELLING

	Authorised Representative		Temperature Limitation
	For in-vitro diagnostic use		Use by/Expiration Date
	Batch Code/Lot number		CAUTION. Consult instructions for use
	Catalogue Number		Manufactured by
	Consult instructions for use		

